

SISTEMI DINAMIČKE FIKSACIJE KUKA



SYSTEMS OF DYNAMIC HIP FIXATION

Narcissus

Implantati i instrumenti za osteosintezu /
Implants and instruments for osteosynthesis

DHS - DCS sistem
/ DHS classic system

DHS ploča / DHS panel

Kataloški broj / Catalog number	Broj otvora / Number of holes	Dužina (mm) / Length (mm)
NA.802	2	47
NA.804	4	79
NA.805	5	95
NA.806	6	111
NA.808	8	143
NA.810	10	175
NA.812	12	207
NA.814	14	239
NA.816	16	271
NA.818	18	303
NA.820	20	335
NA.822	22	367
NA.824	24	399

Materijal: ISO 5832-1 (medicinski čelik)
/ Material for implantation Stainless Steel ISO 5832/1

DHS sistem sastoji se od DHS ploče 135° DHS - DCS klina i kompresivnog zavrtnja. Implantat je dizajniran za fiksaciju preloma proksimalnog femura distalno od preloma. Neophodno je aplicirati 4-6 kortikalnih zavrtnja.

DHS system consists of DHS panel 135°, DHS-DCS screw and a compressive screw. Implant is designed for fixation of proximal femoral fractures distal from the fracture. It is necessary to apply 4-6 cortical screws.



DHS ploča sa kompresivnim žljebovima
135°, 140°, 145°, 150°
/ Compression plate with grooves
135°, 140°, 145°, 150°



Trohanterna T ploča
/Trochanteric plate

Kataloški broj
/Catalog number

NA.P.TT



Materijal: ISO 5832-1 (medicinski čelik)
/ Material for implantation Stainless Steel ISO 5832/1
DHS sistem sa trohanternom T pločom
/DHS system with trochanteric plate

DHS - DCS klin / DHS Locking screw

Kataloški broj / Catalog number	Dužina (mm) / Length (mm)
NA.8050	50
NA.8055	55
NA.8060	60
NA.8065	65
NA.8070	70
NA.8075	75
NA.8080	80
NA.8085	85
NA.8090	90
NA.8095	95
NA.80100	100
NA.80105	105
NA.80110	110
NA.80115	115
NA.80120	120
NA.80125	125



DHS - DCS klin
/ DHS Locking screw

Materijal: ISO 5832-1 (medicinski čelik)
/ Material for implantation Stainless Steel ISO 5832/1

Kompresivni zavrtanj / Compression screw

Kataloški broj / Catalog number	Dužina (mm) / Length (mm)
NA.K.8025	25
NA.K.8030	30



Kompresivni zavrtanj
/ Compression screw

Materijal: ISO 5832-1 (medicinski čelik)
/ Material for implantation Stainless Steel ISO 5832/1

DHS - DCS sistem / DCS system

DCS ploča / DCS panel

Kataloški broj / Catalog number	Broj otvora / No. of holes	Dužina (mm) / Length (mm)
NA.DCS4	4	83
NA.DCS5	5	99
NA.DCS6	6	115
NA.DCS7	7	131
NA.DCS8	8	147
NA.DCS9	9	163
NA.DCS10	10	179
NA.DCS12	12	211
NA.DCS14	14	243
NA.DCS16	16	275
NA.DCS18	18	307
NA.DCS20	20	339
NA.DCS22	22	371
NA.DCS24	24	403

Materijal: ISO 5832-1 (medicinski čelik)
/ Material for implantation Stainless Steel ISO 5832/1

DCS sistem sastoji se od DCS ploče 95° DHS - DCS klina i kompresivnog zavrtnja. Implantat je dizajniran za fiksaciju gornjeg (subtrohanternog) i donjeg (suprakondilarnog i transkondilarnog) okrajka butne kosti.

DCS system consists of DCS panel 95°, DHS-DCS screw and compressive screw. Implant is designed for fixation of upper (subtrochanter) and lower (supracondylar and transcondylar) part of femur.



Pločica sa kompresivnim žljebovima 95°
/ Compression plate with grooves 95°

DKP sistem
/ DKP system

DKP ploča / DKP panel

Kataloški broj / Catalog number	Broj otvora / No. of holes	Dužina (mm) / Length (mm)
NA.402	2	45
NA.404	4	77
NA.405	5	93
NA.406	6	109
NA.408	8	141
NA.410	10	173
NA.412	12	205
NA.413	13	221
NA.414	14	237

DKP sistem se sastoji od DKP ploče i dva glavenovratna zavrtnja. Implant je dizajniran za fiksaciju preloma proksimalnog femura. Distalno od preloma je potrebno aplicirati kortikalne zavrtnje.

DKP system consists of DKP plate and two neck screws. Implant is designed for fixation of proximal femur fractures. It is necessary to apply cortical screws distally from the fracture.

Materijal: ISO 5832-1 (medicinski čelik)
/ Material for implantation Stainless Steel ISO 5832/1



Pločica sa kompresivnim žljebovima
135°, 140°, 145°, 150°
/ Compression plate with grooves
135°, 140°, 145°, 150°

Glavenovratni zavrtanj / Neck screw

Kataloški broj / Catalog number	Dužina (mm) / Length (mm)
NA.4070	70
NA.4075	75
NA.4080	80
NA.4085	85
NA.4090	90
NA.4095	95
NA.40100	100
NA.40105	105
NA.40110	110
NA.40115	115
NA.40120	120
NA.40125	125

Materijal: ISO 5832-1 (medicinski čelik)
/ Material for implantation Stainless Steel ISO 5832/1



Glavenovratni zavrtanj
/ Neck screw